NONINVASIVE DETECTION OF ARTERIAL OCCLUSIVE DISEASE: A THEORETICAL AND MODEL STUDY

L. Stratouly, *† W. Durgin, * F. Anderson Jr., † H. Wheeler† * Mech. Eng., WPI, † Dept. of Surgery, UMASS Med. School

INTRODUCTION - To develop a noninvasive method of detecting arterial occlusive disease, pulse waveforms were recorded at two locations on the lower extremities of young normal volunteers and patients with arteriographically confirmed arterial occlusive disease. Pulses were monitored using impedance plethysmography at the knee and the iliac regions. The frequency spectra of the abnormal iliac waveforms contained 4-5 harmonics as compared to the 2-3 harmonics present in normals'. It was hypothesized that the occurance of high frequencies resulted from pulse wave interactions with diseased portions of the vessel. This paper will present the results of a theoretical and experimental model developed to test this hypothesis.

CLINICAL BACKGROUND - Pulses were monitored at two levels of the extremities in 4 normal volunteers and 13 patients. Information about distal flow in the leg was provided by waveforms obtained when pairs of strip electrodes were placed directly proximal and distal to the knee. Volume pulses from the iliac arteries were acquired using an original electrode array, consisting of two ECG electrodes placed just below the navel and two electrodes placed on the top of the thigh. Figure 1 shows sample pulse waveforms recorded at each location in: a normal volunteer, a patient with severe disease at one level of the limb, and a patient with severe disease at a combination of levels.




Figure 1 - Sample Tracings

The knee pulses of the normal volunteer exhibited a large primary peak and a second, smaller peak. Patient tracings display a damped primary peak and a highly attenuated or absent second wave. This observation may indicate disease and correlates well with the findings of other investigators [1].

The morphology of the iliac pulses reveals some anomalies. The normal volunteer waveforms showed clear primary pulses with solitary secondary pulses. The pulses for the patient with single disease location had many smaller peaks present. This was an unexpected finding. It is hypothesized that these harmonics are the result of pulse interactions with arterial obstsructive disease (AOD).

This observation can be quantified by performing spectral analysis on this signal. Figure 2 shows the power spectrum of a normal volunteer and a patient with AOD. The normal spectrum shows no significant peaks beyond the second harmonic. Five harmonics are present in the diseased case. The goal of the present study is to understand the fundamental basis of this difference in arterial pulse characteristics.




Figure 2 - Power Spectrum Comparison

EXPERIMENTAL ANALYSIS - An experimental model was developed to study wave propagation in a straight compliant tube. silicone rubber tube was used to model the aorta. Experiments modeling a perfect reflecting site, were achieved with a 96 cm. tube having a completely closed end. An oscillating diaphragm provided a sinusoidal input of specified frequency. Flow was the proximal end ο£ the tube at electromagnetic flowmeter and transducers located a distance 15 and 92 cm. away from the diaphragm recorded pressure. pressure readings provided a means of comparing variations in pulse morphology and allowed calculation of wave speed and attenuation.

Experiments were performed using oscillations from 1 to 10 Hz. Flow, pressure and sinusoidal input were monitored with a 4-channel signal processor. The signal processor provided a means of obtaining averaged time and frequency domain data. A plotter, provided a permanent record of results.

Figure 3 depicts the results obtained from the model at a frequency of 8 Hz. All driving frequencies tested displayed similar results, an essentially sinusoidal signal for both pressure and flow. The power spectrum (average of 50 records) had one major peak at the fundamental driving frequency and only minor peaks at the harmonics.




Figure 3: Experimental Results

THEORETICAL MODEL - The theoretical model developed to simulate wave propagation incorporated a frequency response analysis based upon steady oscillatory forced vibrations. The linearized equations for oscillatory flow were solved by a separation of variables technique to determine constitutive equations for flow and pressure head. The functions were represented as the product of spatial and time varying quantities [2]:

Press. Head h' = Re
$$[H(x)e^{i\omega t}]$$
 (1)
Flow q' = Re $[Q(x)e^{i\omega t}]$ (2)

Transfer equations for the complex valued spatial component of flow and pressure head were determined from the particular solutions. The equations were:

$$H(x) = H_{D,U} \cosh \gamma x \pm Z_{c}Q_{D,U} \sinh \gamma x \qquad (3)$$

$$Q(x) = -(H_{D,U}/Z_{c}) \sinh \gamma x \pm Q_{D,U} \cosh \gamma x \qquad (4)$$
where:
$$H_{D,U} = \text{downstream or upstream pressure head}$$

 $Q_{D,U}$ = downstream or upstream flow

γ = propagation constant, a function of: resistance, capacitance, and inertance

Z_c = characteristic impedance, a parameter based on constants of the system

Hydraulic impedance, the ratio of pressure head to flow, was the basis of the third constitutive equation. The transfer equation for impedance was:

The impedance at any point along the tube may also be determined from the reflection coefficient. Reflection in the system is caused by a change in characteristic impedance and is quantified through the complex valued reflection coefficient. An open end tube has a coefficient of (-1.0,0.0), a closed end tube (1.0,0.0), and an infinitely long tube a coefficient of (0.0,0.0). The hydraulic impedance and reflection coefficient are related by:

$$Z(x) = Z_c \left((\Gamma(x) + 1) / (\Gamma(x) - 1) \right) \tag{6}$$

where : Γ = reflection coefficient

STRAIGHT ELASTIC TUBE RESULTS - To solve these equation, boundary conditions were stated at the proximal and distal points of the tube. The proximal end boundary condition specified flow magnitude from the experimental data. reflection end, impedance was specified from the end. Upstream impedance closed а coefficient (5) and upstream pressure using equation calculated With from the impedance and known flow. determined upstream parameters evaluated, the pressure and flow profile along the tube was calculated from eqns. (3) and (4). pressures at 15 and 92 cm. were compared to the values obtained experimentally, Figure 4.




Figure 4 - Comparison of Theory and Experiment

Both models displayed a steady pressure with one resonante peak, at about 5 Hz. The theory and model agree quantitatively with both pressures being of the same order of magnitude. Broader resonance peaks in the experimental model may be due to the uncontrolled source impedance induced by the oscillator and electromagnetic flowmeter.

- The behavior of sinusoidal pulses CONCLUSIONS compliant tube has been demonstrated in straight experimental model study. While theoretical and an amplitude of the pulses was seen to vary with location along as with the source frequency; higher tube as well data, were not similar to those seen in human harmonics, It is concluded that these harmonics are not due demonstrated. to the effect of a complete arterial obstruction. Further are underway to evaluate the contribution of experiments nonlinear elements, including less than complete obstruction, branching and variable compliance.

REFERENCES

[1] Filho, W.S., et. al., "Spectral Analysis of Electrical Impedance Measurements on the Lower Limbs", IEEE Transactions on Biomedical Engineering, Vol., BME-30, No., July, 1983. [2] Streeter, V., Wylie E., Hydraulic Transients, F.E.B. Press, Mich., 1983.